

Herhaalbaarheid spirometrie

Naar welke cijfers kijkt men om te beoordelen of aan de herhaalbaarheid is voldaan?

GOED					FOUT				
	FEV1	FVC	PEF	FEV1/FVC		FEV1	FVC	PEF	FEV1/FVC
Base	0.92	2.29	3.43	40	Base	0.79	2.29	3.43	34
Base	0.95	2.33	3.46	41	Base	0.95	2.33	3.46	41
Base	0.91	2.26	3.09	40	Base	0.78	2.26	3.09	35
Base+	0.95	2.33	3.46	41	Base+	0.95	2.33	3.46	41
Verschil twee hoogste FEV1<150 ml					Verschil twee hoogste FEV1>150 ml				
EN					OF				
	FEV1	FVC	PEF	FEV1/FVC		FEV1	FVC	PEF	FEV1/FVC
Base	0.92	2.29	3.43	40	Base	0.92	2.17	3.43	42
Base	0.95	2.33	3.46	41	Base	0.95	2.33	3.46	41
Base	0.91	2.26	3.09	40	Base	0.91	2.16	3.09	42
Base+	0.95	2.33	3.46	41	Base+	0.95	2.33	3.46	41
Verschil twee hoogste FVC<150 ml					Verschil twee hoogste FVC>150 ml				